



# PERRY JOHNSON LABORATORY ACCREDITATION, INC.

## Certificate of Accreditation

*Perry Johnson Laboratory Accreditation, Inc. has assessed the Laboratory of:*

***American Scale Corporation***  
3540 Bashford Avenue, Louisville, KY 40218

*(Hereinafter called the Organization) and hereby declares that Organization is accredited in accordance with the recognized International Standard:*

**ISO/IEC 17025:2005**

This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory quality management system (as outlined by the joint ISO-ILAC-IAF Communiqué dated January 2009):

***Laboratory and Field Calibration of Weighing Devices***  
*(As detailed in the supplement)*

Accreditation claims for such testing and/or calibration services shall only be made from addresses referenced within this certificate. This Accreditation is granted subject to the system rules governing the Accreditation referred to above, and the Organization hereby covenants with the Accreditation body's duty to observe and comply with the said rules.

For PJLA:

*Initial Accreditation Date:*

January 9, 2004

*Issue Date:*

January 25, 2015

*Expiration Date:*

January 25, 2017

*Accreditation No:*

59225

*Certificate No:*

L15-51

Tracy Szerszen  
President/Operations Manager

Perry Johnson Laboratory  
Accreditation, Inc. (PJLA)  
755 W. Big Beaver, Suite 1325  
Troy, Michigan 48084

*The validity of this certificate is maintained through ongoing assessments based on a continuous accreditation cycle. The validity of this certificate should be confirmed through the PJLA website: [www.pjllabs.com](http://www.pjllabs.com)*



# Certificate of Accreditation: Supplement

## American Scale Corporation

3540 Bashford Avenue, Louisville, KY 40218

Janet Conway Phone: 502-451-5040

Accreditation is granted to the facility to perform the following calibrations:

### Mass Force and Weighing Devices

MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE OR NOMINAL DEVICE SIZE AS APPROPRIATE	CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY ( $\pm$ )	CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED
Lab Balances <sup>o</sup>	0.005 g to 100 g	$(1.00 \times 10^{-4} + 2.73 \times 10^{-6}Wt)g$	Method NIST HB44 NIST IR 6919 ASTM E617 Class 1 Weights
Industrial Scales	1 g to 500 g	$(1.15 \times 10^{-2} + 1.41 \times 10^{-4}Wt)g$	Method NIST HB44 NIST IR 6919 NIST HB 105-1 (Class F) Test Weights
Balances	0.001 lb to 10 lb	$(1.16 \times 10^{-3} + 4.80 \times 10^{-5}Wt)g$	
Bench Scales <sup>o</sup>			
Industrial Scales Bench, Crane, Floor, Livestock, Lift Truck, Tank & Hopper Scales <sup>o</sup>	10 lb to 20 000 lb	$(2.31 + 4.80 \times 10^{-5}Wt)lb$	
Heavy Capacity Industrial Scales Tank & Hopper Scales <sup>o</sup>	20 000 lb to 100 000 lb	$(11.27 + 2.53 \times 10^{-5}Wt)lb$	
Vehicle Scales <sup>o</sup>	1 000 lb to 200 000 lb	$(23.09 + 7.48 \times 10^{-6}Wt)lb$	

1. The CMC (Calibration and Measurement Capability) stated for calibrations included on this scope of accreditation represents the smallest measurement uncertainty attainable by the laboratory when performing a more or less routine calibration of a nearly ideal device under nearly ideal conditions. It is typically expressed at a confidence level of 95 % using a coverage factor  $k$  (usually equal to 2). The actual measurement uncertainty associated with a specific calibration performed by the laboratory will typically be larger than the CMC for the same calibration since capability and performance of the device being calibrated and the conditions related to the calibration may reasonably be expected to deviate from ideal to some degree.
2. The laboratories range of calibration capability for all disciplines for which they are accredited is the interval from the smallest calibrated standard to the largest calibrated standard used in performing the calibration. The low end of this range must be an attainable value for which the laboratory has or has access to the standard referenced. Verification of an indicated value of zero in the absence of a standard is common practice in the procedure for many calibrations but by its definition it does not constitute calibration of zero capacity.
3. The presence of a superscript O means that the laboratory performs calibration of the indicated parameter onsite at customer locations. Example: Outside Micrometer<sup>o</sup> would mean that the laboratory performs this calibration onsite at the customer's location.
4. The term Wt represents weight in pounds or grams (including SI multiple and submultiple units) appropriate to the uncertainty statement.